

101 N. Alloy Drive  
Fenton , MI 48430



PH: 810-714-5811  
FAX: 810-714-5711

CustomerService@lmicorporation.com

Research, Development and Manufacturing of Precision Measuring Systems

Page 1

Cert.# 061623-027

## Certificate of Calibration

**Calibration Performed By:**

LMI CORPORATION  
101 N. ALLOY DR.  
FENTON, MI 48430

**For:**

LMI CORPORATION  
101 N. ALLOY DRIVE  
FENTON MI 48430

**Gage S/N** 5C610227  
**Description** Image Dimension Measuring System  
**Manufacturer** Keyence Corporation  
**Gage Type** Image Dimension Measuring System  
**Unit of Meas.** INCH/METRIC  
**Temperature** 73 F  
**Humidity** 48 %

**Gage ID** LMI CORPORATION - 5C610227  
**Model No.** IM-6225  
**Tol. +**  
**Tol. -**  
**Calibrated By** ALAN BAGGETT  
**As Found Condition** In  
**Calibration Results** Passed  
**Cal. Date** 6/16/2023

**No Cal. Due Date is reported by LMI. This decision is left to customer to best fit their QMS based on freq. of usage**

Test Point Item	Nominal	Tol. +	Tol. -	Before	Deviation	After	Deviation 2	Units
Pass/Fail								

**Findings**

Steel Pin: 4.998mm  
Ceramic Pin: 5.003mm

Ref Standard	Gage S/N	Standard Due Date	Uncert	NIST No
LMI CORPORATION - LMI 104	LMI 104	9/3/2026	6.6E-05	683/283519-13
LMI CORPORATION - LMI 105	LMI 105	9/3/2026	6.5E-05	683/283519-13

It is hereby certified that the above described instrument conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the NIST within the limitations of the Institute Calibration Services or have been derived from accepted values of natural physical constants or have been derived by the ratio type of self calibration techniques. Our calibration system satisfies ISO-9001 and IATF 16949 requirements. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. Measurement Uncertainty is 5.0E-05 An LMI Lab Scope is available upon request.