# A. A. Jansson, Inc.

# **Certificate of Calibration**

2070 Airport Rd Waterford, 48327

Phone No. 248-674-4811 ~ Fax No. 248-674-1234

Company: LINEAR MEASUREMENT INSTRUMENTS

Address: 101 N. ALLOY RD.

FENTON, MI 48430

Contact: ALAN BAGGETT

**Job #:** 85/120

**Calibration Date:** 12/20/2022 **PO/Acct:** 70018512

Page: 1 of 2 Visual Condition: GOOD

Date Received:

Doc #:

Gage Desc.: 60 MM LINEAR GAGE W/ AMPLIFIER Control: 70 985 648F MT 60K 635

005 970

43201

Calibration

Mfg: HEIDENHAIN MT 60K Model: MT 60K GAGE HEAD &

DRO

**Procedure #:** AAJ CM-31.0 **ID#:** 70 985 648F MT 60K 635


005 970

**+ / - Tolerances:** +0.001000 / -0.001000

**Graph Scale:** +0.000050

## AS FOUND DATA

<del>-, .,, .</del>	
Actual	Deviation
+1.000000	+0.000000
+5.000000	+0.000000
+10.000500	+0.000500
+20.000500	+0.000500
+30.000000	+0.000000
+40.000500	+0.000500
+50.000000	+0.000000
+60.000500	+0.000500
Uncertainty	: ± 0.3 μm
	+1.000000 +5.000000 +10.000500 +20.000500 +30.000000 +40.000500 +50.0000000 +60.000500



+ / - Tolerances: 100.000% of +0.000500

**Graph Scale:** +0.000050

### ANVIL FLATNESS

Measurement	Uncertainty:	± 0.20 µm
+0.000000	+0.000500	+0.000500
Nominal	Actual	Deviation

#### 

## Comments:

CLEANED, CALIBRATED AND CERTIFIED DATA IS REPORTED MILLIMETERS

TEMPERATURE: 70.8° F

**HUMIDITY 32.2%** 

.0005 MM RESOLUTION

CERTIFICATE ISSUE DATE: 12/20/2022

NOTE\* NEW GAGE HEAD (MT 60K 635 005 970)

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

The Gage Record for G-3384 could not be found!

96049544: EXTECH TEMPERATURE/HUMIDITY GAGE

5175430036, 5447410003, 3013-0018, 1011-0342-1, 3011-0386 Cal Date / Due Date: 4/6/2022 -- 4/6/2023

160001: 88 PC. METRIC GAGE BLOCK SET

685-O-0000000362-19 Cal Date / Due Date: 12/13/2022 -- 12/13/2023

Gage Status: PASS Due Date: 12/20/2024

# A. A. Jansson, Inc.

## **Certificate of Calibration**

2070 Airport Rd Waterford, 48327

Phone No. 248-674-4811 ~ Fax No. 248-674-1234

**Page #:** 2 of 2

Gage Desc.: 60 MM LINEAR GAGE W/ AMPLIFIER

Control: 7

70 985 648F MT 60K 635

005 970

43201

Mfg: HEIDENHAIN MT 60K

Model:

Doc #:

MT 60K GAGE HEAD &

DRO

Procedure #: AAJ CM-31.0 ID#:

70 985 648F MT 60K 635 005 970

Unless otherwise noted in comments, the following will apply:

- 1. Reported reading were taken at 68° F ± 4° F, Relative humidity less than 50%.
- 2. Compliance statements will be binary, and a simple acceptance rule will be applied, with no consideration

for measurement uncertainty. The tolerance used will be per manufacturers specification.

- 3. This is an endorsed Calibration Certificate (A2LA 1182.01) Valid until 8/31/2024 in accordance with ISO/IEC 17025:2005
- 4. Data is "AS FOUND" and item is "USED"
- 5. The results reported relate only to the items calibtared

This calibration was performed "ONSITE" at customers facility...

Reported measurement uncertainities are based upon a 95% confidence level, k=2. This report shall not be reproduced except in full, without written permission by A. A. Jansson, Inc. Any number of factors may cause the subject of this calibration to drift out of calibration before the recommended interval has expired.

Calibrations performed in accordance with the A. A. Jansson, Inc. Quality Assurance Manual Revision 10, Date: 2020-05-29. Calibration date is the same as "date received"

We certify the equipment used for this calibration traceable to the International System of units (SI) via standards traceable to NIST or other National Metrology Institutes (NMI).

Certified By: Steve Wilt Signature: \_\_\_\_\_Authorized By: \_\_\_\_\_

Brian Lemon - Technicia

This certificate is not valid unless all 2 page(s) are present.

Results of this certificate relate only to the item mentioned in document header.

\*\*END OF DOCUMENT\*\*