

101 N. Alloy Drive  
Fenton , MI 48430



PH: 810-714-5811  
FAX: 810-714-5711

CustomerService@lmicorporation.com

Research, Development and Manufacturing of Precision Measuring Systems

Page 1

Cert.# 031521-001

## Certificate of Calibration

**Calibration Performed By:**

LMI CORPORATION  
101 N. ALLOY DR.  
FENTON, MI 48430

**For:**

LMI CORPORATION  
101 N. ALLOY DRIVE  
FENTON MI 48430

**Gage S/N** LE-060-138  
**Description** Leading Edge Calibration Plate - .060"  
**Manufacturer**  
**Gage Type**  
**Unit of Meas.** INCH  
**Temperature** 68 F  
**Humidity** 60 %

**Gage ID** LMI Corp - LE-060-138  
**Model No.** LMI LG-LE-030  
**Tol. +**  
**Tol. -**  
**Calibrated By** Master Gage & Tool  
**As Found Condition** In  
**Calibration Results** Passed  
**Cal. Date** 11/11/2019

**No Cal. Due Date is reported by LMI. This decision is left to customer to best fit their QMS based on freq. of usage**

Test Point Item	Nominal	Tol. +	Tol. -	Before	Deviation	After	Deviation 2	Units
01 - Thickness	0.0600			0.0623	0.0023	0.0623	0.0023	In

**Findings**

Ref Standard	Gage S/N	Standard Due Date	Uncert	NIST No
LMI CORPORATION - 7.	7 377 845 D	6/13/2024		821/253315

It is hereby certified that the above described instrument conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the NIST within the limitations of the Institute Calibration Services or have been derived from accepted values of natural physical constants or have been derived by the ratio type of self calibration techniques. Our calibration system satisfies ISO-9001 and IATF 16949 requirements. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. Measurement Uncertainty is 5.0E-05 An LMI Lab Scope is available upon request.